

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001555  
11:02 01/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:56  
STERILIZING 3:57  
136.0 C 220kPa 3:57  
136.4 C 225kPa 4:27  
136.4 C 225kPa 4:57  
136.4 C 222kPa 5:27  
136.2 C 223kPa 5:57  
136.3 C 223kPa 6:27  
136.2 C 221kPa 6:57  
136.4 C 225kPa 7:27  
136.4 C 225kPa 7:27  
VENTING 7:27  
AIR DRYING 8:17  
CYCLE COMPLETE 13:35  
-----

Digital Signature #  
33157FDA14517F24  
-----