
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002649
9:52 03/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:51
STERILIZING 3:59
136.5 C 224kPa 3:59
136.4 C 225kPa 4:29
136.3 C 222kPa 4:59
136.4 C 222kPa 5:29
135.9 C 221kPa 5:59
136.2 C 223kPa 6:29
136.4 C 224kPa 6:59
136.4 C 225kPa 7:29
136.4 C 225kPa 7:29
VENTING 7:29
AIR DRYING 8:16
CYCLE COMPLETE 38:16

Digital Signature #
CE57353EFE59BB07
