
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002222
14:09 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:24
136.4 C 226kPa 4:24
136.4 C 225kPa 4:54
136.4 C 223kPa 5:24
136.3 C 225kPa 5:54
136.2 C 222kPa 6:24
136.2 C 223kPa 6:54
136.3 C 225kPa 7:24
136.4 C 225kPa 7:54
136.4 C 225kPa 7:54
VENTING 7:54
AIR DRYING 8:46
CYCLE COMPLETE 8:54

Digital Signature #
998160A97F17A339
