

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3277

Name: RITTER 1  
Sterilizer ID: V2178198  
05/24/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
2:00 191.0 F 1.0  
4:00 205.8 F 6.1  
6:00 218.4 F 11.7  
8:00 227.5 F 7.4  
10:00 241.8 F 13.1  
12:00 241.5 F 11.3  
14:00 240.1 F 10.3  
16:00 254.2 F 17.4  
18:00 266.5 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 29.0  
2:00 272.1 F 28.4  
3:00 271.7 F 28.4  
4:00 272.7 F 29.1  
  
Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 230.3 F 0.0  
10:00 238.1 F 0.0  
15:00 233.2 F 0.0  
20:00 226.6 F 0.0  
25:00 225.4 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:36  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

05/24/2023 02:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_