
Midmark M11 - v1.0.0
Total Cycles: 3989

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2023 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.7 F 0.0
2:00 165.4 F 2.7
4:00 169.1 F 5.9
6:00 199.9 F 10.4
8:00 221.9 F 5.3
10:00 219.6 F 7.7
12:00 230.9 F 12.0
14:00 232.6 F 8.6
16:00 242.3 F 11.8
18:00 244.1 F 12.6
20:00 256.0 F 18.7
22:00 266.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.2
2:00 271.9 F 28.7
3:00 271.2 F 28.5
4:00 272.2 F 29.3

Min 271.1 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 215.8 F 0.0
10:00 225.9 F 0.0
15:00 229.5 F 0.0
20:00 229.2 F 0.0
25:00 227.8 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

11/07/2023 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____