
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000377
13:37 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:14
STERILIZING 3:19
136.0 C 230kPa 3:19
136.3 C 229kPa 3:49
136.1 C 230kPa 4:19
136.3 C 233kPa 4:49
136.4 C 233kPa 5:19
136.3 C 232kPa 5:49
136.3 C 233kPa 6:19
136.3 C 232kPa 6:49
136.3 C 232kPa 6:49
VENTING 6:49
AIR DRYING 7:39
CYCLE COMPLETE 7:46

Digital Signature #
4EF93D688BE683D7
