
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004879
15:04 20/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:22
STERILIZING 3:23
136.4 C 220kPa 3:23
136.3 C 224kPa 3:53
136.4 C 223kPa 4:23
136.4 C 222kPa 4:53
136.2 C 222kPa 5:23
136.2 C 219kPa 5:53
136.4 C 221kPa 6:23
136.1 C 225kPa 6:53
136.1 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:41
CYCLE COMPLETE 10:43

Digital Signature #
45E6161D3060C2BC
