

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000735  
9:33 30/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:30  
STERILIZING 4:36  
136.4 C 224kPa 4:36  
136.3 C 224kPa 5:06  
136.0 C 218kPa 5:36  
136.3 C 221kPa 6:06  
136.3 C 223kPa 6:36  
CYCLE ABORTED 6:40  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 6:40  
-----

Digital Signature #  
92B44FBD30B80902  
-----