

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005654  
10:29 02/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:42  
STERILIZING 3:49  
136.4 C 222kPa 3:49  
136.3 C 224kPa 4:19  
136.4 C 223kPa 4:49  
136.2 C 222kPa 5:19  
136.3 C 224kPa 5:49  
136.1 C 220kPa 6:19  
136.4 C 225kPa 6:49  
136.2 C 220kPa 7:19  
136.2 C 220kPa 7:19  
VENTING 7:19  
AIR DRYING 8:12  
CYCLE COMPLETE 16:01  
-----

Digital Signature #  
89FEBC6D35D3E449  
-----