

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3834

Name: RITTER2  
Sterilizer ID: V2313228  
08/07/2024 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.5 F 0.0  
2:00 135.8 F 1.4  
4:00 142.9 F 3.9  
6:00 174.2 F 7.5  
8:00 191.8 F 11.9  
10:00 215.4 F 5.7  
12:00 227.4 F 10.5  
14:00 231.4 F 7.7  
16:00 240.5 F 12.8  
18:00 240.2 F 11.3  
20:00 252.7 F 17.5  
22:00 263.5 F 24.0  
24:00 272.5 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.6  
1:00 272.3 F 30.4  
2:00 271.7 F 29.6  
3:00 272.1 F 30.4  
4:00 272.0 F 29.7

Min 271.1 F 29.5  
Max 272.9 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 215.1 F 0.0  
10:00 214.2 F 0.0  
15:00 211.4 F 0.0  
20:00 205.9 F 0.0  
25:00 202.0 F 0.0  
30:00 190.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:02:38  
CYCLE COMPLETE

08/07/2024 01:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_