
Midmark M11 - v1.0.5
Total Cycles: 2178

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 171.7 F 3.4
4:00 183.3 F 7.8
6:00 203.6 F 13.3
8:00 223.6 F 7.7
10:00 236.6 F 13.2
12:00 238.8 F 11.2
14:00 238.7 F 10.3
16:00 253.1 F 17.4
18:00 265.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.8
2:00 271.9 F 29.1
3:00 271.6 F 29.3
4:00 272.8 F 29.8

Min 271.3 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 207.3 F 0.0
10:00 210.6 F 0.0
15:00 215.2 F 0.0
20:00 211.3 F 0.0
25:00 210.1 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

03/27/2023 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____