
Midmark M11 - v1.0.5
Total Cycles: 2852

Name: RITTER2
Sterilizer ID: V2313228
09/08/2023 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 177.7 F 2.9
4:00 186.7 F 7.4
6:00 206.9 F 12.5
8:00 219.8 F 7.2
10:00 233.7 F 12.5
12:00 237.0 F 9.9
14:00 236.5 F 9.0
16:00 248.7 F 15.1
18:00 260.5 F 21.7
20:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.3
2:00 272.1 F 29.7
3:00 271.3 F 29.0
4:00 272.4 F 29.6

Min 271.0 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 235.4 F 0.0
10:00 228.1 F 0.0
15:00 229.4 F 0.0
20:00 224.4 F 0.0
25:00 220.0 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

09/08/2023 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____