
Midmark M11 - v1.0.5
Total Cycles: 3672

Name: RITTER2
Sterilizer ID: V2313228
06/06/2024 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING

C383: HEATUP MODE
DOOR OPEN
CHAMBER TEMP 70F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

06/06/2024 08:27 AM