
Midmark M11 - v1.0.5
Total Cycles: 2429

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 192.6 F 4.1
4:00 201.6 F 9.4
6:00 221.0 F 6.9
8:00 230.3 F 10.5
10:00 232.3 F 8.0
12:00 241.6 F 10.8
14:00 247.3 F 14.0
16:00 259.5 F 21.1
18:00 270.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.5
2:00 271.4 F 28.9
3:00 272.5 F 29.7
4:00 271.9 F 29.1

Min 271.3 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 215.6 F 0.0
10:00 234.7 F 0.0
15:00 230.2 F 0.0
20:00 226.4 F 0.0
25:00 229.1 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

05/29/2023 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____