

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3639

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2024 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.8 F 0.0  
2:00 99.5 F 0.1  
4:00 117.7 F 1.2  
6:00 146.2 F 4.3  
8:00 173.9 F 7.7  
10:00 194.0 F 11.6  
12:00 215.8 F 3.7  
14:00 228.5 F 7.3  
16:00 237.9 F 11.4  
18:00 231.7 F 7.7  
20:00 243.7 F 13.0  
22:00 241.6 F 11.4  
24:00 253.3 F 17.3  
26:00 263.6 F 23.5  
28:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.5 F 29.4  
2:00 271.3 F 28.7  
3:00 272.3 F 29.6  
4:00 272.3 F 29.0

Min 271.1 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 214.8 F 0.0  
10:00 187.4 F 0.0  
15:00 179.0 F 0.0  
20:00 176.4 F 0.0  
25:00 173.3 F 0.0  
30:00 169.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 28:17  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:06:37  
CYCLE COMPLETE

05/23/2024 01:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_