
Midmark M11 - v1.0.5
Total Cycles: 882

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 171.8 F 2.2
4:00 172.5 F 5.5
6:00 189.3 F 10.0
8:00 219.6 F 6.5
10:00 226.7 F 10.0
12:00 233.8 F 8.5
14:00 241.0 F 12.4
16:00 240.6 F 11.0
18:00 253.2 F 17.5
20:00 264.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.2 F 29.6
2:00 271.6 F 29.8
3:00 271.9 F 29.4
4:00 271.9 F 30.0

Min 271.2 F 29.2
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 206.6 F 0.0
10:00 218.0 F 0.0
15:00 227.5 F 0.0
20:00 231.0 F 0.0
25:00 224.8 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

03/16/2022 10:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____