
Midmark M11 - v1.0.0
Total Cycles: 2645

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.0 F 0.0
2:00 129.1 F 0.0
4:00 172.8 F 1.0
6:00 169.6 F 4.8
8:00 202.4 F 9.6
10:00 221.5 F 7.0
12:00 227.5 F 7.8
14:00 237.9 F 12.7
16:00 237.2 F 9.7
18:00 237.9 F 8.6
20:00 248.7 F 14.5
22:00 261.1 F 21.5
24:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 271.5 F 28.3
2:00 272.1 F 28.6
3:00 272.2 F 29.0
4:00 271.2 F 28.2

Min 271.1 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 221.3 F 0.0
10:00 234.0 F 0.0
15:00 227.8 F 0.0
20:00 225.4 F 0.0
25:00 226.2 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

12/06/2022 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____