
Midmark M11 - v1.0.5
Total Cycles: 2225

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 178.8 F 3.5
4:00 190.0 F 8.1
6:00 210.5 F 12.3
8:00 224.8 F 9.0
10:00 233.7 F 9.8
12:00 240.6 F 12.6
14:00 241.0 F 11.5
16:00 254.6 F 18.5
18:00 266.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.9
2:00 271.4 F 29.1
3:00 272.3 F 30.1
4:00 272.3 F 29.6

Min 271.2 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 225.3 F 0.0
10:00 225.4 F 0.0
15:00 225.3 F 0.0
20:00 222.6 F 0.0
25:00 224.0 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

04/10/2023 03:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____