
Midmark M11 - v1.0.0
Total Cycles: 906

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2021 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.3 F 0.0
2:00 103.3 F 0.0
4:00 113.8 F 0.1
6:00 143.6 F 0.2
8:00 168.5 F 1.6
10:00 188.8 F 4.1
12:00 221.8 F 6.7
14:00 219.3 F 10.0
16:00 229.8 F 10.8
18:00 230.4 F 8.3
20:00 240.9 F 12.7
22:00 239.6 F 10.0
24:00 238.2 F 9.5
26:00 247.8 F 13.8
28:00 258.4 F 19.6
30:00 268.2 F 25.7
32:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.2 F 28.5
2:00 272.6 F 28.8
3:00 271.9 F 28.7
4:00 271.9 F 28.4

Min 271.5 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 228.2 F 0.0
10:00 217.2 F 0.0
15:00 187.2 F 0.0
20:00 182.5 F 0.0
25:00 180.5 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:34
DRYING: 26:51
TOTAL CYCLE: 01:07:19

06/16/2021 02:07 PM
ACCEPTED: X
REJECTED:
Operator: U01