

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5277

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2025 05:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.3 F 0.0  
2:00 174.4 F 4.5  
4:00 197.0 F 9.9  
6:00 220.9 F 4.7  
8:00 233.5 F 10.3  
10:00 231.5 F 6.8  
12:00 245.2 F 12.5  
14:00 244.8 F 12.4  
16:00 258.8 F 19.8  
18:00 270.9 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.3 F 28.3  
2:00 272.7 F 28.9  
3:00 271.8 F 27.9  
4:00 272.7 F 28.7  
  
Min 271.5 F 27.8  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 210.7 F 0.0  
10:00 205.9 F 0.0  
15:00 215.9 F 0.0  
20:00 230.5 F 0.0  
25:00 244.1 F 0.0  
30:00 232.3 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 18:41  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

01/24/2025 06:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_