
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000415
7:15 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:36
STERILIZING 3:47
136.0 C 229kPa 3:47
135.8 C 229kPa 4:17
136.4 C 233kPa 4:47
136.2 C 230kPa 5:17
136.2 C 231kPa 5:47
136.4 C 233kPa 6:17
136.4 C 232kPa 6:47
136.4 C 232kPa 7:17
136.4 C 232kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 8:22

Digital Signature #
B637D6F8D88DDC66
