
Midmark M11 - v1.0.0
Total Cycles: 952

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2021 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.6 F 0.0
2:00 121.4 F 0.0
4:00 137.3 F 0.0
6:00 153.3 F 0.4
8:00 175.8 F 2.9
10:00 197.4 F 5.6
12:00 208.9 F 9.2
14:00 221.1 F 13.0
16:00 225.3 F 6.8
18:00 235.9 F 11.0
20:00 235.2 F 8.1
22:00 244.8 F 12.9
24:00 242.5 F 11.2
26:00 253.7 F 16.9
28:00 263.7 F 22.9
30:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.4 F 28.8
2:00 272.2 F 28.4
3:00 271.3 F 28.0
4:00 272.1 F 28.8

Min 271.3 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 219.4 F 0.0
10:00 226.3 F 0.0
15:00 229.6 F 0.0
20:00 232.8 F 0.0
25:00 224.1 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

06/29/2021 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____