
Midmark M11 - v1.0.0
Total Cycles: 4537

Name: RITTER 1
Sterilizer ID: V2178198
05/09/2024 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 78.3 F 0.0
4:00 125.4 F 0.5
6:00 126.9 F 2.7
8:00 151.9 F 5.3
10:00 174.6 F 8.4
12:00 202.1 F 12.2
14:00 214.9 F 3.7
16:00 219.6 F 7.2
18:00 229.5 F 10.9
20:00 229.1 F 6.8
22:00 239.4 F 10.8
24:00 234.1 F 7.7
26:00 248.2 F 14.2
28:00 259.3 F 20.3
30:00 269.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.5
2:00 271.8 F 28.5
3:00 272.1 F 28.3
4:00 272.2 F 28.8

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 213.0 F 0.0
10:00 192.5 F 0.0
15:00 193.1 F 0.0
20:00 200.3 F 0.0
25:00 221.3 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

05/09/2024 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____