
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001627
12:50 26/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:44
STERILIZING 3:56
136.5 C 221kPa 3:56
136.4 C 225kPa 4:26
136.4 C 225kPa 4:56
136.4 C 224kPa 5:26
136.4 C 222kPa 5:56
136.3 C 223kPa 6:26
136.3 C 223kPa 6:56
136.2 C 222kPa 7:26
136.2 C 222kPa 7:26
VENTING 7:26
AIR DRYING 8:18
CYCLE COMPLETE 8:31

Digital Signature #
FB2D3F1917004EBB
