

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1737

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2022 07:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 79.3 F 0.0  
4:00 109.8 F 0.0  
6:00 115.0 F 0.1  
8:00 145.2 F 1.6  
10:00 170.7 F 4.2  
12:00 192.8 F 7.6  
14:00 211.1 F 11.4  
16:00 222.0 F 4.4  
18:00 229.6 F 8.8  
20:00 239.1 F 13.0  
22:00 236.6 F 9.6  
24:00 238.3 F 9.2  
26:00 246.5 F 13.3  
28:00 258.1 F 19.5  
30:00 268.7 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.2 F 28.4  
2:00 272.5 F 28.7  
3:00 272.8 F 29.1  
4:00 272.0 F 28.9

Min 271.6 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 217.1 F 0.0  
10:00 206.3 F 0.0  
15:00 225.2 F 0.0  
20:00 220.2 F 0.0  
25:00 213.6 F 0.0  
30:00 193.4 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:46  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:09:37  
CYCLE COMPLETE

03/24/2022 08:10 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02