

-----  
Midmark M11 - v1.0.0  
Total Cycles: 793

Name: RITTER 1  
Sterilizer ID: V2178198  
04/21/2021 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
2:00 191.2 F 4.3  
4:00 205.1 F 9.9  
6:00 222.3 F 5.4  
8:00 234.0 F 11.0  
10:00 234.5 F 8.7  
12:00 238.8 F 9.5  
14:00 249.1 F 14.8  
16:00 261.9 F 22.1  
18:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.3 F 28.7  
2:00 272.2 F 28.9  
3:00 272.1 F 28.5  
4:00 272.5 F 29.1  
  
Min 271.6 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 207.4 F 0.0  
10:00 213.1 F 0.0  
15:00 216.3 F 0.0  
20:00 195.4 F 0.0  
25:00 193.6 F 0.0  
30:00 189.3 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 18:15  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:27  
CYCLE COMPLETE

04/21/2021 12:51 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01