
Midmark M11 - v1.0.5
Total Cycles: 1010

Name: RITTER2
Sterilizer ID: V2313228
04/20/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.2 F 0.0
2:00 177.2 F 0.3
4:00 197.9 F 4.6
6:00 216.8 F 9.9
8:00 226.0 F 7.8
10:00 237.8 F 12.5
12:00 239.6 F 11.1
14:00 239.9 F 10.6
16:00 254.3 F 18.0
18:00 266.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.8
3:00 272.6 F 29.6
4:00 272.0 F 29.0

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 208.9 F 0.0
10:00 218.9 F 0.0
15:00 223.6 F 0.0
20:00 225.3 F 0.0
25:00 226.4 F 0.0
30:00 225.1 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

04/20/2022 10:02 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____