
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003642
8:36 11/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:43
136.4 C 224kPa 3:43
135.9 C 221kPa 4:13
136.1 C 223kPa 4:43
136.4 C 224kPa 5:13
136.2 C 222kPa 5:43
136.5 C 225kPa 6:13
136.3 C 222kPa 6:43
136.4 C 222kPa 7:13
136.4 C 222kPa 7:13
VENTING 7:14
AIR DRYING 8:35
CYCLE COMPLETE 12:26

Digital Signature #
9F9A95D41296560F
