

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.1uS / 1.3ppm  
CYCLE NUMBER 000034  
15:13 10/08/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 3:39  
STERILIZING 4:47  
136.4 C 227kPa 4:47  
136.4 C 228kPa 5:17  
136.4 C 227kPa 5:47  
136.3 C 225kPa 6:17  
136.4 C 227kPa 6:47  
136.4 C 225kPa 7:17  
136.4 C 227kPa 7:47  
136.3 C 226kPa 8:17  
VENTING 8:17  
AIR DRYING 9:09  
CYCLE COMPLETE 39:09  
-----

Digital Signature #  
323BE3AC3BC994E7  
-----