
Midmark M11 - v1.0.0
Total Cycles: 4449

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 134.1 F 1.6
4:00 146.4 F 4.6
6:00 186.6 F 7.9
8:00 204.9 F 11.9
10:00 220.0 F 5.2
12:00 228.1 F 9.2
14:00 235.6 F 13.5
16:00 234.8 F 9.7
18:00 240.4 F 9.7
20:00 245.1 F 13.1
22:00 256.2 F 19.0
24:00 266.5 F 25.3
26:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.6 F 29.5
2:00 272.2 F 29.6
3:00 272.4 F 29.3
4:00 272.5 F 29.8

Min 271.8 F 29.1
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 209.2 F 0.0
10:00 210.7 F 0.0
15:00 207.3 F 0.0
20:00 216.2 F 0.0
25:00 230.7 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:22
CYCLE COMPLETE

04/11/2024 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____