
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005457
8:47 02/05/2023

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	2:40
PRESSURIZING	4:47
STERILIZING	5:42
136.0 C 220kPa	5:42
136.4 C 224kPa	6:13
136.3 C 222kPa	6:43
136.3 C 224kPa	7:13
136.3 C 221kPa	7:43
136.4 C 224kPa	8:13
136.4 C 224kPa	8:43
136.3 C 223kPa	9:13
136.4 C 224kPa	9:43
136.3 C 224kPa	10:13
136.4 C 222kPa	10:43
136.4 C 221kPa	11:13
136.2 C 220kPa	11:43
136.3 C 222kPa	12:13
136.2 C 220kPa	12:43
136.4 C 223kPa	13:13
136.3 C 223kPa	13:43
136.3 C 225kPa	14:13
136.4 C 224kPa	14:43
136.4 C 224kPa	15:13
136.3 C 223kPa	15:43
136.3 C 223kPa	15:43
VENTING	15:43
AIR DRYING	16:33
CYCLE COMPLETE	16:39

Digital Signature #

7BD9A19AED020442
