

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000146  
10:31 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:17  
PRESSURIZING 4:17  
STERILIZING 5:50  
136.4 C 225kPa 5:50  
136.1 C 223kPa 6:20  
136.1 C 223kPa 6:50  
136.3 C 225kPa 7:20  
136.0 C 223kPa 7:50  
136.4 C 225kPa 8:20  
136.2 C 224kPa 8:50  
136.2 C 222kPa 9:20  
136.2 C 222kPa 9:20  
VENTING 9:20  
AIR DRYING 10:25  
CYCLE COMPLETE 10:28  
-----

Digital Signature #  
0BF0DD4AA91CB6C9  
-----