
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006668
13:37 23/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:33
STERILIZING 3:36
136.4 C 222kPa 3:36
136.4 C 223kPa 4:06
136.3 C 222kPa 4:36
136.1 C 219kPa 5:06
136.2 C 218kPa 5:36
136.4 C 221kPa 6:06
136.4 C 224kPa 6:36
136.0 C 221kPa 7:06
136.0 C 221kPa 7:06
VENTING 7:06
AIR DRYING 7:56
CYCLE COMPLETE 8:03

Digital Signature #
1B31E1EFEB419E72
