
Midmark M11 - v1.0.5
Total Cycles: 1942

Name: RITTER2
Sterilizer ID: V2313228
01/18/2023 08:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 79.8 F 0.0
4:00 96.8 F 0.0
6:00 111.6 F 0.0
8:00 161.9 F 0.1
10:00 187.4 F 0.5
12:00 205.5 F 4.1
14:00 218.4 F 7.4
16:00 231.0 F 11.2
18:00 229.7 F 6.9
20:00 240.8 F 12.1
22:00 238.5 F 10.2
24:00 236.9 F 8.8
26:00 250.0 F 15.6
28:00 261.4 F 22.1
30:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.3
2:00 271.7 F 29.1
3:00 272.4 F 29.2
4:00 271.7 F 29.0

Min 271.5 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 227.5 F 0.0
10:00 231.8 F 0.0
15:00 229.7 F 0.0
20:00 229.7 F 0.0
25:00 226.6 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

01/18/2023 09:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____