
Midmark M11 - v1.0.5
Total Cycles: 2733

Name: RITTER2
Sterilizer ID: V2313228
08/04/2023 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 69.3 F 0.0
4:00 75.9 F 0.0
6:00 98.4 F 0.0
8:00 150.8 F 0.0
10:00 177.7 F 0.0
12:00 194.4 F 2.1
14:00 206.3 F 4.9
16:00 215.5 F 7.7
18:00 222.6 F 11.0
20:00 227.2 F 7.5
22:00 235.7 F 10.3
24:00 236.6 F 9.4
26:00 244.5 F 13.1
28:00 243.1 F 12.0
30:00 254.3 F 17.8
32:00 264.3 F 23.8
34:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.4
2:00 271.6 F 28.5
3:00 272.6 F 29.5
4:00 272.2 F 28.9

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 207.7 F 0.0
10:00 207.9 F 0.0
15:00 206.5 F 0.0
20:00 201.6 F 0.0
25:00 198.7 F 0.0
30:00 193.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 34:12
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:12:31
CYCLE COMPLETE

08/04/2023 10:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____