
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000331
11:19 18/12/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 5:36
STERILIZING 6:41
136.5 C 233kPa 6:41
136.3 C 232kPa 7:11
136.4 C 232kPa 7:41
CYCLE ABORTED 7:59
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:59

Digital Signature #
2A8743EDC4E4A1E4
