
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002486
9:32 07/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:02
STERILIZING 3:11
136.5 C 220kPa 3:11
136.2 C 223kPa 3:41
136.0 C 221kPa 4:11
136.0 C 220kPa 4:41
136.2 C 223kPa 5:11
136.3 C 224kPa 5:41
136.3 C 225kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:31
CYCLE COMPLETE 37:31

Digital Signature #
B2F5B508EF1E4D28
