
Midmark M11 - v1.0.5
Total Cycles: 3422

Name: RITTER2
Sterilizer ID: V2313228
02/29/2024 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.1 F 0.0
2:00 141.4 F 1.6
4:00 159.6 F 4.6
6:00 182.8 F 9.0
8:00 205.9 F 8.8
10:00 221.3 F 7.7
12:00 233.5 F 12.7
14:00 235.9 F 10.0
16:00 236.0 F 8.6
18:00 249.5 F 15.4
20:00 261.7 F 22.4
22:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.8
3:00 272.6 F 29.7
4:00 272.0 F 29.1

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 206.0 F 0.0
10:00 199.7 F 0.0
15:00 199.6 F 0.0
20:00 196.3 F 0.0
25:00 194.2 F 0.0
30:00 192.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

02/29/2024 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____