
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002161
7:53 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:56
136.4 C 224kPa 3:56
136.3 C 223kPa 4:26
136.4 C 225kPa 4:56
136.4 C 224kPa 5:26
136.4 C 224kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:56
136.4 C 224kPa 7:26
136.4 C 224kPa 7:26
VENTING 7:26
AIR DRYING 8:21
CYCLE COMPLETE 16:42

Digital Signature #
F3D6C75661E527EC
