
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000240
9:59 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:45
STERILIZING 5:17
136.0 C 228kPa 5:17
136.4 C 232kPa 5:47
136.4 C 233kPa 6:17
136.4 C 233kPa 6:47
136.3 C 232kPa 7:17
136.3 C 232kPa 7:47
136.3 C 232kPa 8:17
136.3 C 233kPa 8:47
136.3 C 233kPa 8:47
VENTING 8:47
AIR DRYING 9:50
CYCLE COMPLETE 11:51

Digital Signature #
12DBC8B27472B46F
