
Midmark M11 - v1.0.0
Total Cycles: 2360

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.8 F 0.0
2:00 170.1 F 3.4
4:00 179.3 F 7.6
6:00 199.9 F 12.6
8:00 218.7 F 6.4
10:00 231.3 F 11.3
12:00 232.4 F 7.7
14:00 243.5 F 13.5
16:00 241.8 F 11.6
18:00 253.9 F 17.8
20:00 264.7 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.7
2:00 272.6 F 29.3
3:00 271.7 F 28.8
4:00 271.9 F 29.3

Min 271.4 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 207.4 F 0.0
10:00 221.8 F 0.0
15:00 225.4 F 0.0
20:00 233.1 F 0.0
25:00 230.7 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 21:60
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

09/22/2022 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____