
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006946
7:42 15/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:28
STERILIZING 4:32
136.4 C 222kPa 4:32
136.3 C 223kPa 5:02
136.4 C 224kPa 5:32
136.2 C 218kPa 6:02
136.1 C 217kPa 6:32
136.3 C 222kPa 7:02
136.0 C 216kPa 7:32
136.3 C 222kPa 8:02
136.3 C 222kPa 8:02
VENTING 8:02
AIR DRYING 8:52
CYCLE COMPLETE 8:58

Digital Signature #
848A31CBA8598E15
