

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006353  
8:12 26/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:54  
STERILIZING 2:48  
136.0 C 220kPa 2:48  
136.3 C 218kPa 3:18  
136.4 C 218kPa 3:48  
136.1 C 222kPa 4:18  
136.3 C 221kPa 4:48  
135.7 C 219kPa 5:18  
136.4 C 222kPa 5:48  
136.4 C 222kPa 6:18  
136.4 C 222kPa 6:18  
VENTING 6:18  
AIR DRYING 7:08  
CYCLE COMPLETE 7:21

-----  
Digital Signature #  
8D5D3020313B4D85  
-----