
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002255
14:51 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:38
PRESSURIZING 0:48
STERILIZING 1:59
136.4 C 224kPa 1:59
136.1 C 222kPa 2:29
136.1 C 223kPa 2:59
136.4 C 225kPa 3:29
136.2 C 223kPa 3:59
136.4 C 224kPa 4:29
135.8 C 220kPa 4:59
136.4 C 225kPa 5:29
136.4 C 225kPa 5:29
VENTING 5:29
AIR DRYING 6:21
CYCLE COMPLETE 9:02

Digital Signature #
AB6F760E7ACF464F
