
Midmark M11 - v1.0.5
Total Cycles: 2375

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 191.3 F 3.9
4:00 206.8 F 9.4
6:00 224.4 F 7.9
8:00 234.5 F 11.6
10:00 235.5 F 9.8
12:00 236.1 F 9.0
14:00 249.8 F 15.8
16:00 261.9 F 22.8
18:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.1 F 29.4
2:00 271.3 F 29.3
3:00 272.0 F 29.4
4:00 271.1 F 29.1

Min 271.1 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 214.0 F 0.0
10:00 235.3 F 0.0
15:00 234.7 F 0.0
20:00 228.1 F 0.0
25:00 224.6 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

05/16/2023 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____