
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002898
11:21 29/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:13
STERILIZING 3:09
136.0 C 220kPa 3:09
135.9 C 220kPa 3:39
136.4 C 222kPa 4:09
136.4 C 222kPa 4:39
136.3 C 224kPa 5:09
136.4 C 221kPa 5:39
136.3 C 223kPa 6:09
136.4 C 223kPa 6:39
136.4 C 223kPa 6:39
VENTING 6:39
AIR DRYING 7:29
CYCLE COMPLETE 7:33

Digital Signature #
A919EE7F62922638
