

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3133

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2023 12:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 167.7 F 2.5  
4:00 178.6 F 6.4  
6:00 199.4 F 10.8  
8:00 216.7 F 5.0  
10:00 225.1 F 10.1  
12:00 232.2 F 8.2  
14:00 242.8 F 13.1  
16:00 242.5 F 12.2  
18:00 255.1 F 18.7  
20:00 266.2 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.8  
2:00 271.7 F 29.0  
3:00 271.9 F 29.7  
4:00 272.6 F 29.6

Min 271.2 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 211.4 F 0.0  
10:00 215.2 F 0.0  
15:00 209.2 F 0.0  
20:00 205.4 F 0.0  
25:00 205.2 F 0.0  
30:00 206.2 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:39  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:00:30  
CYCLE COMPLETE

11/16/2023 01:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_