
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.6uS / 1.0ppm
CYCLE NUMBER 001322
12:33 27/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 4:06
136.0 C 220kPa 4:06
136.4 C 224kPa 4:36
136.4 C 225kPa 5:06
136.4 C 224kPa 5:36
136.4 C 225kPa 6:06
136.2 C 222kPa 6:36
136.0 C 221kPa 7:06
136.1 C 221kPa 7:36
136.1 C 221kPa 7:36
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 23:28

Digital Signature #
BCC2D5B1FB9FD67C
