

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3162

Name: RITTER 1  
Sterilizer ID: V2178198  
04/25/2023 01:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.5 F 0.0  
2:00 156.1 F 0.0  
4:00 190.8 F 0.0  
6:00 209.4 F 3.7  
8:00 223.9 F 8.2  
10:00 236.3 F 13.0  
12:00 235.9 F 9.7  
14:00 239.5 F 8.8  
16:00 243.7 F 11.6  
18:00 247.2 F 13.6  
20:00 260.1 F 20.7  
22:00 271.7 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.5 F 28.6  
2:00 271.7 F 28.3  
3:00 272.4 F 28.6  
4:00 271.5 F 28.1

Min 271.5 F 28.0  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 220.7 F 0.0  
10:00 236.2 F 0.0  
15:00 232.4 F 0.0  
20:00 227.3 F 0.0  
25:00 225.1 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:33  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:44  
CYCLE COMPLETE

04/25/2023 02:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_