

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006146  
9:00 17/08/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:35  
STERILIZING 3:29  
136.0 C 220kPa 3:29  
136.4 C 224kPa 3:59  
136.4 C 223kPa 4:29  
136.3 C 222kPa 4:59  
136.4 C 224kPa 5:29  
136.1 C 220kPa 5:59  
136.4 C 223kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 224kPa 6:59  
VENTING 6:59  
AIR DRYING 7:52  
CYCLE COMPLETE 7:55

-----  
Digital Signature #  
E2A0A8FE4B645103  
-----